




# Test Certificate

Part 1 Liter Vessel  
Part Number 14322-01  
Serial Number 101791 - 101832



		USP / Ph.Eur. Tolerances		
Dimensions	Vessel	Inside Ø a	98.5 to 101.5 mm	98 – 106 mm
		Height b	200 mm	160 – 210 mm
		With hemispherical bottom		
Material	Borosilicate glass			

The accuracy of the above mentioned part dimensions have been examined. All relevant measurements are within tolerance and the factory acceptance test has passed. The dimensions are in conformance with USP and Ph. Eur. Pharmacopoeia requirements (at date of approval, see below).

Aesch, 2025-01-21

Tayyib Özüdogru  
SOTAX Quality Management  
SOTAX AG | Nordring 1 | 4147 Aesch | Switzerland